

GC Sample Introduction

an eLearning course



About the Instructor

Dr. Gerard Sharp has more than 25 years experience in chromatography. This experience has ranged from the stringent requirements of working in a key forensic drug lab (Melbourne) to the production stresses associated with a high sample throughput environmental lab (as Chief Chemist). Gerard spent 15 months in a major government research lab in Japan developing methods for low level sampling of halomethanes for analysis by bench top and high resolution GC-MS. Other experience includes roles as a senior analyst for a leading paint company and a multinational chemical manufacturer. More recently he has been employed as the Global Technical Manager for a major GC consumables supplier so understands manufacture and specification of supplies and how this relates to performance. Gerard has carried out technical training throughout Australia, Europe, the USA and Asia and appreciates language and cultural barriers to learning. Gerard has a B.Sc (Hons) in synthetic organic chemistry from Monash University and a PhD in Analytical Chemistry from the University of New South Wales, Australia.

Ardent Scientific Pty Ltd makes no representations or warranties of any kind, express or implied as regard to the material in this document; including but not limited to the implied warranties of merchantability and fitness for a particular purpose, except as otherwise provided under applicable laws.

Neither Ardent Scientific Pty Ltd, nor its affiliates, directors, officers, employees, agents, contractors, successors or assigns will be liable for any damages whatsoever arising out of, or in any way related to, the use of information from this document. This limitation applies to direct, indirect, consequential, special, punitive or other damages you or others may suffer, as well as damages for loss of profits, business interruption or the loss of data or information.

This document and its contents are provided to you on "as is" basis, the document may contain errors, faults and inaccuracies and may not be complete and current. Ardent Scientific shall not be liable for errors contained herein or for incidental, or consequential damages in connection with the furnishing, performance, or use of this material.

The information contained in this document is subject to change without notice.

This training manual and the Ardent Scientific web site and its contents are subject to copyright. The material copyright is owned by Ardent Scientific Pty Ltd.

No part of this document and no part of the contents of the web site may be reproduced, modified, translated or distributed without the prior written consent of Ardent Scientific.

This document includes registered trademarks, and other trademarks that are otherwise protected by law. Except as expressly authorized, the use or misuse of any of these trademarks is strictly prohibited.

Included Trademarks and Registered Trademarks are Trademarks of their respective owners.

© 2012 by Ardent Scientific Pty Ltd

www.ardentscientific.com.au

All rights reserved

Table Of Contents

ABOUT THE INSTRUCTOR	ii
SAMPLE INTRODUCTION - OVERVIEW	1
WHAT YOU WILL LEARN	2
ROLE OF THE GC INLET	3
THE AIRPORT ANALOGY	4
THE SPLIT / SPLITLESS INLET	5
LEARNING ACTIVITY	6
SUMMARY	7
SPLIT INJECTION	9
WHAT YOU WILL LEARN	10
ANATOMY OF A SPLIT / SPLITLESS INLET	11
SEPTUM PURGE FLOW	12
SPLIT FLOW AND COLUMN FLOW	13
SPLITTING THE SAMPLE	14
CONTROLLING THE SPLITTING	15
SPLITTING THE TOURISTS	16
WHY SPLIT – COLUMN CAPACITY	17
WHY SPLIT – FAST SAMPLE TRANSFER	18
LEARNING ACTIVITY 1	19
LEARNING ACTIVITY 2	20
SUMMARY	21
SPLITLESS INJECTION	22
WHAT YOU WILL LEARN	23
SPLITLESS FLOWS	24
SPLITLESS SAMPLING	25
PEAK FOCUSING – COLD TRAPPING	26
PEAK FOCUSING – SOLVENT EFFECT	27
PULSED SPLITLESS	28
OPENING THE SPLIT VALVE	29
LEARNING ACTIVITY 1	30
LEARNING ACTIVITY 2	31
SUMMARY	32